Insulin Pump Back-Up Plan

If your pump is not working, call Medtronic/Tandem/OmniPod using phone number on back of pump. The representative may be able to help fix your pump & order a replacement pump to be sent to you. Until you get a new pump, use the following back-up plan.

1.	Check blood glucose and correct using your fast-acting insulin (Humalog/Novolog/Apidra/Fiasp) and a syringe.
2.	Take back-up long-acting insulin (Lantus/Levemir/Basaglar/Toujeo/Tresiba) units every 24 hours. Write down the time:
•	
3.	Use syringe and fast-acting insulin (Humalog/Novolog/Apidra/Fiasp) to give
	meal-time bolus using carbohydrate ratio 1: and correction/sensitivity
	factor 1: for BG over You may wish to download an insu
	dosing app on your smartphone.
4.	Check your blood sugar levels more frequently.
5.	When you have a replacement pump, you can choose to set it up at the same
	time you gave your long-acting insulin (see step #2). Or, set up your pump as
	soon as you get it. Use a temporary basal rate of 0% to last until the time noted in step #2.
6.	Use your previous insulin pump settings to set up your new pump. You can refe

to your last insulin pump download report or call our office for these settings.